

Will Sub-Nanosecond Lasers Replace Nanosecond Lasers for Tattoo Removal?

Boris Cencic, Zdenko Vizintin, Matjaz Lukac
Fotona d.d., Ljubljana, Slovenia

SUMMARY

With their short nanosecond pulse durations, Q-switched lasers have revolutionized the treatment of tattoos [1-3]. Nevertheless, the mechanisms for clinically observed tattoo clearance are still not well understood. Two mechanisms have been proposed: a) a thermal effect where tattoo particles are heated to sufficiently high temperatures to cause chemical changes within the particles and the surrounding cells; and b) a mechanical fragmentation of the particles due to extremely fast changes in the particles' temperature [1-4]. Theoretically, both mechanisms should be more effective towards shorter pulse durations. It is for this reason that sub-nanosecond, 750-900 picoseconds (e.g., 0.75 - 0.9 nanoseconds) tattoo-removal lasers [5, 9], have recently attracted considerable interest. However, as measurements of the dependence of the ablation threshold fluence on pulse duration in a wide range of pulse durations (as short as 0.1 ps) have demonstrated, the pulse duration is of importance only when the irradiated tissue is basically transparent to the laser light [6]. For highly absorbing tissues, such as tattooed skin, the ablation threshold fluence was found to be completely independent of the laser pulse duration, even when pulse durations were shortened from the nanosecond to sub-picosecond (0.1 ps) range [6]. The question thus arises whether the shortening of laser pulses will have any significant effect on tattoo removal efficacy, especially since the pulse duration of the current "picosecond" devices are only slightly shorter than one nanosecond.

In this contribution, we report on the preliminary results of a study of the influence of laser pulse duration on tattoo removal.

When tattooed skin is treated with a laser pulse fluence above a certain treatment threshold, plasma formation takes place and gas bubbles form around the tattoo pigment [7, 8]. This transformation is observed clinically as a whitening or blanching of the treated skin, signifying that the tattoo pigment reacted with the treatment light.

We measured the dependence of the plasma

formation fluence, as observed when skin whitening occurred, on the treatment laser pulse duration. A laboratory Nd:YAG laser set-up was used which was able to generate pulses of the following durations (Table 1):

Table 1

Nd:YAG laser pulse duration			
50 ns	7 ns	5 ns	2 ns

Laser fluences where skin whitening occurred were measured on tattooed pig and human skin. For both, pig and human tattooed skin, and for the above wide range of tested pulse durations, the plasma formation threshold changed only slightly, from 0.8 to 1.00 J/cm², even though the pulse duration varied by a factor of 25 (from 2 to 50 nanoseconds). It is highly unlikely that a shortening of the pulse duration by another factor of 2.5 (to obtain sub-nanosecond pulses) would result in any further significant change. In agreement with previous studies which have showed the threshold for plasma formation in highly absorbing tissues to be insensitive to pulse duration, our results indicate that reducing the pulse duration into the sub-nanosecond range will not contribute significantly to the thermal mechanism for tattoo removal.

Another mechanism which could possibly improve with sub-nanosecond (750-900 ps) pulses, is the fracturing of tattoo particles under increased mechanical stress. However, as has been shown by previously published studies, tattoo particle fragmentation does not occur even when 20-times shorter (e.g., 35 ps) pulses are used [4]. A conclusion by the study was [4] that temperature-induced changes, rather than particle fragmentation, are responsible for tattoo clearing.

It is also worth noting that picosecond pulses of a sufficiently high fluence are difficult to generate, and that consequently the picosecond lasers are capable of delivering fluences above plasma formation threshold only at small spot sizes. These small spot sizes result not only in procedures being slow, but also in

unacceptable scattering losses, so that tissue penetration and treatment efficacy are compromised [4]. For example, current sub-nanosecond lasers are capable of generating a maximum fluence of 6.4 J/cm² only at the 2 mm spot size, while at the 3 mm spot size the maximum fluence falls down to only 2.8 J/cm² [5]. This is to be compared with the top-of-the-line Q-switched Nd:YAG lasers with 5 ns pulse durations, which can generate fluences of 12.7 J/cm² at 3 mm, 8.2 J/cm² at 5 mm, 5.7 J/cm² at 6 mm, and 3.2 J/cm² at 8 mm [7].

Clinical experience has also not confirmed that sub-nanosecond lasers are effective independently of whether the laser wavelength is matched with the color of the pigment. In a recent study with a sub-nanosecond alexandrite laser [9], no clearance of a red colored tattoo was obtained after four treatment sessions.

Another challenge with picosecond lasers is the optical breakdown in the air above the treated skin that occurs more readily at shorter pulse durations [4]. This might limit the usefulness of picosecond lasers because plasma would consume energy intended for treating tattoo particles. The picosecond lasers are also more complex devices and might be more difficult to maintain and service.

In conclusion, the published literature and our preliminary results seem to indicate that current “picosecond” lasers are not expected to have a significantly better tattoo clearance effect in comparison with the “gold standard” Q-switched nanosecond tattoo lasers. Taking into account also the current technology limitations of picosecond lasers, it is our opinion that top-of-the-line Q-switched nanosecond lasers will remain the devices of choice for tattoo removal.

REFERENCES

1. S. Choudhary, M. L. Elsaie, A. Leiva, and K. Nouri, "Lasers for tattoo removal: a review," *Lasers in Medical Science* 25, 619-627 (2010).
2. M. A. Adatto, "Laser tattoo removal: Benefits and Caveats," *Med. Laser Appl.* 19, 175-185 (2004).
3. E.F. Bernstein, "Laser treatment of tattoos," *Clinics in Dermatology*, Volume 24(1), Issue 1, Pages 43-55 (2006).
4. Ross EV, Naseef G, Lin C, Kelly M, Michaud N, Flotte TJ, Raythen J, Andersen R, Comparison of responses of tattoos to picosecond and nanosecond Q-switched Neodymium:YAG lasers, *Arch Dermatol*, Vol 134, Feb 1998; 167-171.
5. Sierra R, Mirkov M, Impact of pulse duration from nanoseconds to picoseconds on the thermal and mechanical effects during laser interaction with tattoo particles, *Lasers in Surg Med* 45, Issue Supplement 25 (March 2013).
6. Oraevsky AA, Da Silva LB, Rubenchik AM, Feit MD, Glinsky ME, Perry MD, Mammini BM, Small W, Stuart BC, Plasma mediated ablation of biological tissues with nanosecond-to-

7. femtosecond laser pulses: relative role of linear and nonlinear absorption, *IEEE J Select Top Q Electr*, Vol 2, No. 4, 1996; 801-809.
7. Cencic, B., Grad, L., Mozina, J., Jezersek, M., Optodynamic monitoring of laser tattoo removal. *J. Biomed. Opt.*, 17:047003-047006, 2012.
8. Cencic B, Lukac M, Marincek M, Vizintin Z, High fluence, high beam quality Q-switched Nd:YAG laser with Optoflex delivery system for treating benign pigmented lesions, *LA&HA J. Laser and Health Academy*, Vol. 2010, No.2; 9-18. www.laserandhealth.com.
9. Robinson DM, Saedi N, Petrell K, Arndt KA, Dover JS, Confirmatory study of picosecond 755nm alexandrite laser, *Lasers in Surg Med* 45, Issue Supplement 25 (March 2013).

The intent of this Laser and Health Academy publication is to facilitate an exchange of information on the views, research results, and clinical experiences within the medical laser community. The contents of this publication are the sole responsibility of the authors and may not in any circumstances be regarded as official product information by medical equipment manufacturers. When in doubt, please check with the manufacturers about whether a specific product or application has been approved or cleared to be marketed and sold in your country.